



DHMH

Maryland Department of Health and Mental Hygiene

Office of Health Care Quality

Spring Grove Center, Bland Bryant Building

55 Wade Avenue, Catonsville, Maryland 21228-4663

Larry Hogan, Governor - Boyd K. Rutherford, Lt. Governor - Dennis R. Schrader, Secretary

Analyzer Pre-field Evaluation & Coefficient of Variation Report

Instrument: Type: _____

Serial Number: _____

Date completed: _____

Evaluation Findings

Level 1 Lot # _____ Exp. Date _____

Mean concentration: _____ mg/dl

Standard Deviation: _____ mg/dl

Coefficient of variation: _____ %

Performance standard: C.V. <20% Yes _____ No _____

Level 2 Lot # _____ Exp. Date _____

Mean concentration: _____ mg/dl

Standard Deviation: _____ mg/dl

Coefficient of variation: _____ %

Performance standard: C.V. <20% Yes _____ No _____

_____ Acceptable for screening

_____ Unacceptable for screening

_____ Other (specify): _____

Performed by: (Analyst) _____

Reviewed by: (Director) _____